



# 防爆合格证

CONFORMITY CERTIFICATE OF EXPLOSION-PROOF

证号  
Certificate No. CE23.6920X

产品名称  
Name of Product 多变量差压式质量流量计

型号及规格  
Type of Product MMF DN MA UD

防爆标志  
Marking Ex db II C T6 Gb

技术文件  
Technical Documents Q/020MKY04-2022

图号  
Drawing No. MMF-0-01

备注  
Note (s)

1. 工作电压: (5~36) VDC, 功率: ≤3W。
2. 使用环境温度:  $-40^{\circ}\text{C} \leq T_a \leq +70^{\circ}\text{C}$ 。
3. 特殊使用条件:  
须配套与此产品防爆等级适用的并取得强制性认证的引入装置。
4. 关联检验报告: J2316920。

经对上述产品图样及技术文件的审查和样品的检验, 其符合以下标准

By verifying the drawings and technical documents and checking samples, the product complies with the following standards

GB/T 3836.1-2021 GB/T 3836.2-2021

发给: 迈科洛 (广州) 仪表有限公司

Issued to:

本证失效日期: 2028-04-20

Date of Expire:

发证日期: 2023-04-20

Date of Issue:

印章  
Seal



主任  
Director

中海油天津化工研究设计院有限公司  
(国家防爆产品质量检验检测中心(天津))  
(石油和化学工业电气产品防爆质量监督检验中心)  
National Explosion-proof Product Quality Inspection and Testing Center(TianJin)

注: 本证仅对与批准的审查认可文件和送检样品一致的产品有效。

Note: This certificate is only valid for the documents approved and for the products that are in accord with samples tested and verified.

中心地址: 中国天津市丁字沽三号路 85 号

Center Add: No.85 No.3 Road DingZiGu Tianjin China

E-mail: pcec@pcec.com.cn

邮政编码: 300131

Post code: 300131

http://www.pcec.com.cn

电话/传真: 022-26651066 /26689116

Tel/ Fax: 022-26651066/26689116